

Declaration of conformity: CryoPen - CryoProbe XP

D9.08.00

Rev: 01

1. Manufacturer: H&O Equipments NV/SA

Rue des Journaliers 1 7822 Ghislenghien

Belgium

2. Products:

Cryogenic Surgical instruments

3. Specified product:

Description	Article Number	
CryoPen XP	S-HO-CEPE-01	
CryoProbe XP	S-HO-CEPR-01	
Applicator for CryoPen XP and CryoProbe XP	S-HO-CCXP-MA-005/006/007/008	
Cartridges 16g for Cryo XP starter sets	S-HO-16g-BOX	
Cartridges 8g for CryoPen XP	S-HO-NOCX-XX-S24/-12-S24	
Cartridges 8g for CryoProbe XP	S-HO-NOCX-XX-S24-PR/-12-S24-PR	
Cartridges 16g for CryoPen XP	S-HO-NOCX-XL-S06/-12-S06	
Cartridges 16g for CryoProbe XP	S-HO-NOCX-PRXL-S06/-PR-12-S06	

We declare under our sole responsibility that those products are in conformity with the essential requirements.

Document Nr.	Title	Date of emission	Risk Classification
93/42/EEC	Directive medical devices	14/06/1993	Class IIa (MDD, rule9)
DORS/98-282	Medical devices regulations	16/08/2012	Class II (MDR, rule9)

- 4. Additional information: (procedure of conformity, Notified Body, CE certificate, etc.)
 - Procedure of conformity for CE marking: MDD, annexes II (excluding section 4)
 - Notified Body: SGS House Noorderlaan 87 2030 Antwerp Belgium, number CE 1639, certificate Nr. BE19/819943764
 - MDSAP/ISO 13485:2016 SGS certificate BE20/819943916

Additional information for CryoProbe | XP (for USA or Canada)

- Caution: Federal Law restricts this device to sale by or on the order of a physician or licensed practitioner.
- Canadian License nr: 89787

Ghislenghien, Belgium

Date: 19/01/2023

Erik Hermans

Chief Executive Officer