

DECLARATION OF CONFORMITY

(in accordance with ISO/IEC 17050-1)

We declare, under our sole responsibility, that the product listed below conforms to the provisions of:

- the European Council Directive 93/42/EEC of 14 June 1993 concerning medical devices
- the Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS)

| | |
|-----------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------|
| Manufacturer's Name and Business Address: | Welch Allyn, Inc. 4341 State Street Road Skaneateles Falls, NY 13153 U.S.A. |
| EC REP | Regulatory Affairs Representative Welch Allyn Limited Navan Business Park Dublin Road Navan, County Meath Republic of Ireland |
| Product Name ^{1,3} : | Illuminator and Trans Illuminator |
| REF _{1,3} | 901025 – Illuminator, Hand Held |
| # _{1,3} | 23857, 26030, 26035, 26530, 26535, 26538, 27000, 27050, 28100, 41100, 41101, 43300 |
| Radio equipment ² : | Not applicable, no radio |
| Object of the declaration ² : | Not applicable, no radio |
| Accessories and components ² : | Not applicable, no radio |
| Medical Device Conformity Assessment Route Annex ¹ : | VII |
| Medical Device Classification ¹ : | I |
| Medical Device Classification Rules ¹ : | 1 & 12 |
| GMDN Code and Term ¹ : | 36761 Skin transilluminator, line-powered 12276 Light Examination |

¹ applicable to the medical devices directive, 93/42/EEC

² applicable to the radio equipment directive, 2014/53/EU

³ applicable to the RoHS directive, 2011/65/EU

| UMDNS Code and Term ¹ : | 14130 Transilluminators | |
|-----------------------------------------------------------------------------------------------------------|-------------------------|------------------------------------------------------------------------------------------------------------------------------------------|
| Standards Applied (Standards are applicable to the medical device directive, unless otherwise indicated): | Number | Title |
| | EN 50581 ³ | Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances |
| | EN 60601-1 | Medical Electrical Equipment – Part 1: General Requirements for Basic Safety and Essential Performance |
| | EN 60601-1-1 | Medical Electrical Equipment – General Requirements for Safety – Collateral Standard: Safety requirements for Medical Electrical Systems |

Authorised Signatory:



 Fiona Butler, Manager Regulatory Affairs
 {EU Authorised Representative}

2019-03-21

 Date

Navan

 Place of Issue

¹ applicable to the medical devices directive, 93/42/EEC
² applicable to the radio equipment directive, 2014/53/EU
³ applicable to the RoHS directive, 2011/65/EU