

**Declaration of Conformity****Manufacturer**

Tyromotion GmbH  
Bahnhofgürtel 59, A-8020 Graz, Austria  
SRN (Single Registration Number): AT-MF-000016168

**Product Specification**

Article Number and Product Name	(11310) MYRO®
UDI-DI	9120077180069
Intended Purpose	The MYRO® therapy system is generally used for neurological rehabilitation of the upper extremities.
Basic UDI-DI	912007718MYROE4
Including Software	TyroS Software Version 6.5.x
Components	MYRO® Therapydesk MYRO® Objects (Coin, Ball, Cup, Grip, Force, Blank)
Classification according to Medical Device Regulation (EU) 2017/745, Annex VIII	Class I according to Rule 1 and 13

**Conformity Assessment**

Conformity Assessment Procedure according to Certificate	Medical Device Regulation (EU) 2017/745, Annex IX, Chap. I This certificates is valid for devices manufactured in the manufacturing period given, only.
Valid for	12. Dec. 2024 - 05. Jan. 2028

This Declaration of Conformity is issued under the sole responsibility of the manufacturer.

The products described above, as delivered, comply with the requirements of Regulation (EU) 2017/745 (MDR), the Machinery Regulation (EU) 2023/1230, the Directive 2014/53/EU (Radio Equipment Directive) and the Austrian Medical Devices Act BGBl. I No. 122/2021 as amended.

This declaration is supported by the certification of the Quality Management System according to EN ISO 13485 by DQS Medizinprodukte GmbH. The products are provided with CE and UKCA marking.

**tyromotion**

TYROMOTION GMBH  
Bahnhofgürtel 59, 8020 Graz, Austria  
Tel: +43 (0) 316 908 909 / Mail: office@tyromotion.com  
ATU63250555 / FN261052v  
www.tyromotion.com

Sabine Seereiner  
Head of QM & RA

Graz, 12. Dec. 2024