

CERTIFICAZIONI/CERTIFICATES:

EN ISO 9001
EN ISO 13485
EN ISO 14001
MDSAP 13485:2016
FDA Establishment Registration Number: 3006846316

**EU DECLARATION OF CONFORMITY
ACCORDING TO REGULATION MDR 745/2017, ARTICLE 19**

**MEDICAL DEVICE FAMILY:
PERIPHERAL IV CATHETERS and ACCESSORIES**

Medical device Manufacturer identification.	
Name, registered trade name	DELTA MED SPA
SRN as referred to in Article 31	IT-MF-000027962
Registered office address	Via Guido Rossa N° 20, Viadana (MN), CAP 46019, Italy
Production site	Via Guido Rossa N° 20, Viadana (MN), CAP 46019, Italy

The present EU Declaration of conformity is issued by Delta Med SpA under its full and sole responsibility, for all the medical devices listed in this document.

All the medical devices listed in this EU Declaration of conformity are referred to the BASIC UDI-DI reported in the table enclosed to this declaration:

Basic UDI DI, reported in this declaration of conformity, is defined according to Part C of Annex VI of EU Regulation 745/2017.

The present EU Declaration of conformity covers all the medical devices listed in Table 1 enclosed to this document which are:

- Device medical device family group: **Peripheral IV Catheters and accessories**
- Intended purpose.

Intended purpose of the accessories

Obturator: Accessory intended to be used in combination only and exclusively with Delta Med single and dual entry IV catheters. This accessory allows complete occlusion of the catheter tube and catheter body in order to ensure closure and suspension of administration.

The risk class of the device was determined according to the rules set out in Annex VIII. In particular Peripheral IV Catheters and accessories are all Class IIa medical devices.

The devices Peripheral IV Catheters and accessories covered by the present declaration are in conformity with EU Regulation 2017/745 requirements.

Peripheral IV Catheters and accessories are in compliance with the following standards used and in relation to which conformity is declared:

CERTIFICAZIONI/CERTIFICATES:

EN ISO 9001

EN ISO 13485

EN ISO 14001

MDSAP 13485:2016

FDA Establishment Registration Number: 3006846316

- EN ISO 13485:2016/A11.2021 - Harmonized standard - Medical devices - Quality management systems - Requirements for regulatory purposes (ISO 13485:2016).
- EN ISO 14971:2019/A11.2021 - Medical devices - Application of risk management to medical devices
- EN ISO 10993-12:2021. Biological evaluation of medical devices - Part 12: Sample preparation and reference materials
- EN ISO 10993-3:2014. Biological evaluation of medical devices - Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity
- EN ISO 10993-4:2017. Biological evaluation of medical devices - Part 4: Selection of tests for interactions with blood
- EN ISO 10993-5:2009. Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity
- EN ISO 10993-6:2016. Biological evaluation of medical devices - Part 6: Tests for local effects after implantation.
- EN ISO 10993-10:2013. Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization.
- EN ISO 10993-11:2018. Biological evaluation of medical devices - Part 11: Tests for systemic toxicity.
- EN ISO 10555-1: 2013/A1.2017. Intravascular catheters - Sterile and single-use catheters - Part 1: General requirements - Amendment 1
- EN ISO 10555-5: 2013. Intravascular catheters - Sterile and single-use catheters - Part 5: Over-needle peripheral catheters
- ISO 14972: 1998. Sterile obturators for single use with over-needle peripheral intravascular catheters
- EN ISO 23908: 2011. Sharps injury protection - Requirements and test methods - Sharps protection features for single-use hypodermic needles, introducers for catheters and needles used for blood sampling.
- EN ISO 9626: 2016. Stainless steel needle tubing for the manufacture of medical devices - Requirements and test methods
- EN ISO 80369-7:2017. Small-bore connectors for liquids and gases in healthcare applications - Part 7: Connectors for intravascular or hypodermic applications
- EN ISO 11135:2014 - Sterilization of health care products – Ethylene Oxide-Part 1 – Requirements for development, validation and routine control of the sterilization process for medical devices
- EN ISO 11607-1:2020- Packaging for terminally sterilized medical devices - Part 1: Requirements for materials, sterile barrier systems and packaging systems
- EN ISO 11607-2:2020 - Packaging for terminally sterilized medical devices - Part 2: Validation requirements for forming, sealing and assembly processes
- EN ISO 11737-1:2018 - Sterilization of health care products - Microbiological methods Determination of a population of microorganisms on products
- EN ISO 11737-2:2020 - Sterilization of medical devices - Microbiological methods - Part 2: Tests of sterility performed in the definition, validation and maintenance of a sterilization process (ISO 11737-2:2009)
- EN ISO 11138-1:2017 - Sterilization of health care products - Biological indicators - Part 1: General requirements (ISO 11138-1:2017)
- EN ISO 11138-2:2017 - Sterilization of health care products - Biological indicators - Part 2: Biological indicators for ethylene oxide sterilization processes (ISO 11138-2:2017)
- EN ISO 15223-1:2021 - Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements.
- EN IEC 62366-1:2015 – Medical device. Part 1 – Application of usability engineering to medical device
- EN ISO 20417:2021. Medical device. Information to be supplied by the manufacturer

CERTIFICAZIONI/CERTIFICATES:

EN ISO 9001
EN ISO 13485
EN ISO 14001
MDSAP 13485:2016
FDA Establishment Registration Number: 3006846316

Notified body identification.	
Notified body	TÜV SÜD Product Service GmbH
Address:	Zertifizierstelle – Ridlerstrasse 65, D - 80339 München
ID Number	CE0123
CE Certificate N°	G20 026056 0032 Rev.00
Issuing date of EC Certificate	2023-02-02
Expiring date of EC Certificate	2028-02-01
Conformity assessment procedure	According to Annex XI, part A

Place and date of issue of the declaration

Place: Viadana, (MN) Date:2023-06-26

Name and function of the person who signed

Signature:



Name: Olga Raschi

Position: QA&RA Manager, PRRC

This declaration of conformity is signed by the above mentioned function, on behalf of:

Name: Gabriele Giovanelli

Position: CEO, Delta Med SpA

CERTIFICAZIONI/CERTIFICATES:

EN ISO 9001
EN ISO 13485
EN ISO 14001
MDSAP 13485:2016
FDA Establishment Registration Number: 3006846316

Medical devices list of the present EU Declaration of conformity.

DEVICE CODE	DESCRIPTION (BRAND NAME DEVICE)	BASIC UDI-DI
SO+DM2419PD	Obt. Terumo Dual entry 24G 19mm	8032248400007P
SO+DM2225PD	Obt. Terumo Dual entry 22G 25mm	8032248400007P
SO+DM2032PD	Obt. Terumo Dual entry 20G 32mm	8032248400007P
SO+DM1845PD	Obt. Terumo Dual entry 18G 45mm	8032248400007P
SO+DM1745PD	Obt. Terumo Dual entry 17G 45mm	8032248400007P
SO+DM1645PD	Obt. Terumo Dual entry 16G 45mm	8032248400007P
SO+DM1445PD	Obt. Terumo Dual entry 14G 45mm	8032248400007P
SO+DM1832PD	Obt. Terumo Dual entry 18G 32mm	8032248400007P
SO+DM2419WD	Obt. Terumo Single entry 24G 19mm	8032248400007P
SO+DM2225WD	Obt. Terumo Single entry 22G 25mm	8032248400007P
SO+DM2032WD	Obt. Terumo Single entry 20G 32mm	8032248400007P
SO+DM1845WD	Obt. Terumo Single entry 18G 45mm	8032248400007P
SO+DM1745WD	Obt. Terumo Single entry 17G 45mm	8032248400007P
SO+DM1645WD	Obt. Terumo Single entry 16G 45mm	8032248400007P
SO+DM1445WD	Obt. Terumo Single entry 14G 45mm	8032248400007P
SO+DM1832WD	Obt. Terumo Single entry 18G 32mm	8032248400007P