



*Declaration of Conformity*

We, 3M Company  
hereby declare under our sole responsibility  
that the CE marked products,

3M™ Attest™ Steam Biological Indicator Incubator (56°C), 118  
3M™ Attest™ EO Biological Indicator Incubator (37°C), 129

to which this declaration relates are in accordance with the

Electro-Magnetic Compatibility Directive 2014/30/EU,  
EN 61326-1 (2013); IEC 61326-1 (2012)

and the

Low Voltage Directive 2014/35/EU,  
IEC 61010-1 (2010)  
IEC 61010-2-010 (2014)

3M Health Care Business self-declares conformity with Directive 2011/65/EU of the European Parliament  
and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances  
in electrical and electronic equipment and compliance to the requirements of EN 50581:2012.

EU Representative Address

3M Deutschland GmbH  
Health Care Business  
Carl-Shurz-Str. 1  
41453 Neuss  
Germany

Signature: \_\_\_\_\_

Dianne Gibbs  
3M Health Care  
Division Regulatory Affairs Manager  
Medical Solutions Division

Date: \_\_\_\_\_

4 May 2020