



**DECLARATION OF CONFORMITY  
TO COUNCIL DIRECTIVE 93/42/EEC  
CONCERNING MEDICAL DEVICES**

<b>MANUFACTURER:</b>	<b>CONTEC MEDICAL SYSTEMS CO., LTD</b> No.112 Qinhuang West Street, Economic & Technical Development Zone, Qinhuangdao, Hebei Province, PEOPLE'S REPUBLIC OF CHINA
<b>MEDICAL DEVICE:</b>	Pulse Oximeter, CMS60D
<b>CLASSIFICATION - ANNEX IX:</b>	Class II b, Rule 10
<b>CONFORMITY ASSESSMENT ROUTE:</b>	Annex II excluding chapter 4
WE, ( CONTEC MEDICAL SYSTEMS CO., LTD ) HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES; INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE.	
STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.	
<b>NOTIFIED BODY:</b>	TÜV SÜD PRODUCT SERVICE GMBH RIDLERSTR 65, D-80339 M NCHEN, GERMANY
<b>IDENTIFICATION NUMBER:</b>	 0123
<b>(EC) CERTIFICATE(S):</b>	<u>G1 050972 0050 Rev.04</u>
<b>EUROPEAN REPRESENTATIVE:</b>	Shanghai International Holding Corp. GmbH(Europe) Eiffestrasse 80, 20537 Hamburg Germany

**START OF CE-MARKING:** 2009-07-23 (Date or Lot or serial number)

<b>PLACE, DATE OF DECLARATION:</b>	QINHUANGDAO, 2020-06-18
<b>SIGNATURE:</b>	 _____ President

# DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

## Appendix: list of (harmonised - EN) standards

No.	Serial Number	Title and Description
1	IEC 60601-1:2012	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance
2	IEC60601-1-2:2014	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests
3	EN 60601-1-6:2010 (IEC 60601-1-6:2010)	Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability
4	EN 60601-1-8:2007 (IEC 60601-1-8:2006)	Medical electrical equipment - Part 1-8: General requirements for basic safety essential performance- Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems
5	ISO 80601-2-61:2011	Medical electrical equipment —Part 2-61:Particular requirements for basic safety and essential performance of pulse oximeter equipment
6	EN 60601-1-11:2010 (IEC 60601-1-11:2010)	Medical electrical equipment –Part 1-11: General requirements for basic safety and essential performance –Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment
7	EN 62366:2008 (IEC 62366:2007)	Medical devices - Application of usability engineering to medical devices
8	EN 62304:2006 (IEC 62304:2006)	Medical device software - Software life-cycle processes