

**EUROPEAN MEDICAL DEVICE REGULATION****Declaration of Conformity**

As Legal Manufacturer, we

3M Company
Single Registration Number (TBD)
2510 Conway Ave. St. Paul, MN 55144 USA

hereby declare under our sole responsibility that the following CE marked devices


Trade Name	3M Red Dot™ Monitoring Electrode with Foam Tape and Sticky Gel 3M Red Dot™ Radiolucent Monitoring Electrode with Foam Tape, Sticky Gel and Abrader Pad 3M Red Dot™ Monitoring Electrode with Foam Tape 3M Red Dot™ Monitoring Electrode with 4 mm Adapter 3M Red Dot™ Radiolucent Monitoring Electrode with Foam Tape
Intended Purpose	Electrocardiograph (ECG) electrode
Reference	2560, 2560-3 & 2560-5 2570, 2570-3 & 2570-5 2228 2228BA 2244
Basic UDI-DI	06082238401010000000042AA

are classified per rule 1 of Annex VIII of the Medical Device Regulation (EU) 2017/745, as Class I devices in accordance with all applicable provisions of the REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL concerning medical devices.

3M Company self-declares conformity with Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment, as amended per (EU) 2015/863, and compliance to the requirements of EN IEC 63000:2018.

EU Authorized Representative:

3M Deutschland GmbH
Health Care Business
Single Registration Number (TBD)
Carl-Schurz-Str. 1
41453 Neuss, Germany



Dianne Gibbs
Regulatory Affairs Director
3M Company

11 August 2021

Date

3M and Red Dot are trademarks of 3M.