

Declaration of Conformity

| Manufacturer | ArjoHuntleigh AB Hans Michelsensgatan 10 211 20 Malmö, Sweden |
|----------------------------------|--|
| Single Registration Number | SE-MF-00000696 |
| Declaration | ArjoHuntleigh AB as the manufacturer of the following medical devices, takes sole responsibility and declares conformity with the applicable provisions of Medical Device Regulation (EU) 2017/745 concerning medical devices, by Annex IX. |
| Device Family Name | D920/D930/FD1/FD3 |
| Intended Purpose | The product is intended to monitor fetal heart rate. |
| Basic UDI-DI | 5060693520358WM |
| Risk Class and Rule | Class IIa, Rule 10 |
| Additional Information | Manufactured and distributed on behalf of ArjoHuntleigh AB by: Huntleigh Healthcare Ltd 35 Portmanmoor Road Cardiff CF24 5HN United Kingdom Also complies with the following EU Legislation: RoHS Directive 2011/65/EU WEEE Directive 2012/19/EU |
| Notified Body Name and Number | BSI Group The Netherlands B.V. Number: 2797 CE Certificate Number MDR 718928 |

| | 1 |
|-----------------------------------|-----------------|
| Title: QRE Compliance Director Si | gnature: Um |
| Name: Steve Monks Da | ate: 19/12/2023 |

On behalf of ArjoHuntleigh AB: Place: Cardiff