## **DECLARATION OF CONFORMITY**



(in accordance with ISO/IEC 17050-1)

Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

We declare, under our sole responsibility, that the product listed below conforms to the provisions of:

- European Council Directive 93/42/EEC of 14 June 1993 concerning medical devices, and
- Directive 2014/53/EU of the European Parliament and of the Council of 16 April 2014 on the harmonization of the laws of the Member States relating to the making available on the market of radio equipment, and
- Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment as amended by Commission Delegated Directive (EU) 2015/863 of 31 March 2015 (RoHS3).

| Document Number 8001630                                  | 2 Version  | K   |  |  |  |                      |  |
|--|--|---|--|--|--|----------------------|--|
| Product Name   | ProBP 3400   | Series  |  |  |  |                      |  |
| Manufacturer's Name and<br>Business Address              | 4341 State   | Welch Allyn, Inc.<br>4341 State Street Road<br>Skaneateles Falls, NY 13153<br>USA |  |  |  | SRN: US-MF-000013394 |  |
| EC Certificates<br>Declaration of Conformity<br>Validity | EC Certificate 314505 MR2<br>Expiry Date: 2024-05-26   |   |  |  |  |                      |  |
| EC REP   | Welch Allyn Limited<br>Navan Business Park, Dublin Road<br>Navan Co. Meath<br>C15 AW22 Ireland |   |  | S  | RN: IE-AR-000                                | 0000768              |  |
| REF #  | 34XFHT-B<br>34XFWT-B<br>34XFHT-2<br>34XXHT-4   | 34XXHT-B<br>34XXWT-B<br>34BXWT-2<br>34XFST-4                                      | 34XXHT-B<br>34XXHT-2<br>34XXWT-2<br>34XFWT-4 | 34XFST-B<br>34XFWT-2<br>34XFST-2<br>34XXST-4 | 34XXST-B<br>34XXST-2<br>34XFHT-4<br>34XXWT-4 |                      |  |
| Radio equipment  | 901055, DIGITAL BLOOD PRESSURE DEVICE Laird Tech, Model BTM411                                 |   |  |  |  |                      |  |

Object of the declaration



Medical Device Conformity Assessment Route Annex Annex II





(in accordance with ISO/IEC 17050-1)

Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

| Medical Device Classification         | lla  |  |  |
|---------------------------------------|--|--|--|
| Medical Device Classification<br>Rule | 10   |  |  |
| Standards                             | See Appendix A   |  |  |
| GMDN Code and Term                    | 45617 – Automatic-inflation electronic sphygmomanometer, portable, arm/wrist                       |  |  |
| UMDNS Code and Term                   | 16173 - Sphygmomanometers, Electronic, Automatic   |  |  |
| Notified Body                         | DQS Medizinprodukte GmbH, August-Schanz-Str.21, 60433 Frankfurt am Main Notified Body Number: 0297 |  |  |



## **DECLARATION OF CONFORMITY**

(in accordance with ISO/IEC 17050-1)

Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

**Authorised Signatory** 

Joshua Kim

Senior Manager

Global Regulatory Affairs

2<u>622 · 04 · 1</u>0

Skaneateles Falls NY, USA Place of Issue

## DECLARATION OF CONFORMITY





Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

Appendix A: Standards and Common Specifications

| Standards Applied       | Number             | Version/Date of<br>Issue | Title   |
|-------------------------|--------------------|--------------------------|---|
| Directive 93/42/EEC     | EN 60601-1         | 2013                     | Medical electrical equipment - Part 1: General requirements for basic safety and essential performance  |
|                         | EN 60601-1-2       | 2015                     | Medical Electrical Equipment, Part 2: Collateral Standard: Electromagnetic Compatibility: Requirements and Test   |
|                         | EN ISO 81060-<br>1 | 2012                     | Non-Invasive Sphygmomanometers - Part 1:<br>Requirements and Test Methods for Non-Automated<br>Measurement Type   |
|                         | EN ISO 81060-<br>2 | 2014                     | Non-Invasive Sphygmomanometers - Part 2: Clinical Validation of Automated Measurement Type  |
|                         | EN 80601-2-<br>30  | 2018                     | Medical Electrical Equipment - Part 2-30: Particular<br>Requirements for The Basic Safety and Essential<br>Performance of Automated Non-Invasive<br>Sphygmomanometers   |
|                         | EN 62366-1         | 2015                     | Medical devices Part 1: Application of usability engineering to medical devices   |
|                         | EN ISO 10993-      | 2010                     | Biological evaluation of medical devices - Part 1:<br>Evaluation and testing  |
|                         | EN 62304           | 2006 check               | Medical Device Software - Software Life Cycle<br>Processes  |
|                         | EN 62353           | 2008                     | Medical Electrical Equipment - Recurrent Test and Test<br>After Repair of Medical Electrical Equipment  |
| Directive<br>2014/53/EU | EN 62311           | 2008 Check               | Assessment of electronic and electrical equipment related to human exposure restrictions for electromagnetic fields (0 Hz – 300 GHz)  |
|                         | EN 301 489-1       | 2019-03 check            | ElectroMagnetic Compatibility (EMC) standard for radio equipment and services; Part 1: Common technical requirements; Harmonised Standard covering the essential requirements of article 3.1(b) of Directive 2014/53/EU and the essential requirements of article 6 of Directive 2014/30/EU |





(in accordance with ISO/IEC 17050-1)

Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

| Standards Applied                          | Number            | Version/Date of Issue | Title  |
|--|-------------------|-----------------------|--|
|  | EN 301 489-<br>17 | 2017-03 check         | ElectroMagnetic Compatibility (EMC) standard for radio equipment and services; Part 17: Specific conditions for Broadband Data Transmission Systems; Harmonised Standard covering the essential requirements of article 3.1(b) of Directive 2014/53/EU   |
|  | EN 300 328        | 2016-11 check         | Electromagnetic compatibility and Radio spectrum Matters (ERM); Wideband transmission systems; Data transmission equipment operating in the 2,4 GHz ISM band and using wide band modulation techniques; Harmonized EN covering the essential requirements of article 3.2 of Directive 2014/53/EU |
|  | EN 301 893        | 2017-03 Check         | 5 GHz RLAN; Harmonised Standard covering the essential requirements of article 3.2 of Directive 2014/53/EU   |
| Directive<br>2011/65/EU<br>+ (EU) 2015/863 | EN IEC 63000      | 2018                  | Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances   |