


Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

We declare, under our sole responsibility, that the product listed below conforms to the provisions of:

- Regulation 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices

Document Number 80016679	Version M	
Product Name	Flexiport Cuff (Reuse)	
Manufacturer's Name and Business Address	Welch Allyn, Inc. 4341 State Street Road Skaneateles Falls, NY 13153 USA	SRN: US-MF-000013394
Declaration of Conformity Validity	ISO 13485 #314505 MP2016 Expiry Date: 2022-12-08	
	Welch Allyn Limited, Navan Business Park, Dublin Road, Navan, Co. Meath, C15 AW22 Ireland	SRN: IE-AR-000000768

Object of the declaration



Flexiport Cuff (Reusable)

Intended Purpose Welch Allyn Pediatric through Adult Blood Pressure cuffs are non-invasive blood pressure cuffs intended for use in conjunction with non-automated and automated sphygmomanometers to determine blood pressure in pediatric through adult patients.

Medical Device Conformity Assessment Route Annex	Annex II and Annex III
Medical Device Classification	Class I
Medical Device Classification Rule	Rules 1
Standards	Refer to Appendix A

Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

901043 BLOOD PRESSURE CUFF, REUSABLE

REF	#
------------	----------

REUSE-06	REUSE-07	REUSE-08	REUSE-09	REUSE-10	REUSE-11
REUSE-06-1HP	REUSE-07-1HP	REUSE-08-1HP	REUSE-09-1HP	REUSE-10-1HP	REUSE-11-1HP
REUSE-06-1MQ	REUSE-07-1MQ	REUSE-08-1MQ	REUSE-09-1MQ	REUSE-10-1MQ	REUSE-11-1MQ
REUSE-06-1TP	REUSE-07-1TP	REUSE-08-1TP	REUSE-09-1TP	REUSE-10-1TP	REUSE-11-1TP
REUSE-06-1TPE	REUSE-07-1TPE	REUSE-08-1TPE	REUSE-09-1TPE	REUSE-10-1TPE	REUSE-11-1TPE
	REUSE-07-2BV	REUSE-08-2BV	REUSE-09-2BV	REUSE-10-2BV	REUSE-11-2BV
	REUSE-07-2BVC	REUSE-08-2BVC	REUSE-09-2BVC	REUSE-10-2BVC	REUSE-11-2BVC
	REUSE-07-2MF	REUSE-08-2MF	REUSE-09-2MF	REUSE-10-2MF	REUSE-11-2MF
	REUSE-07-2MQ	REUSE-08-2MQ	REUSE-09-2MQ	REUSE-10-2MQ	REUSE-11-2MQ
	REUSE-07-2TP	REUSE-08-2TP	REUSE-09-2TP	REUSE-10-2TP	REUSE-11-2TP
	REUSE-07-2TPE	REUSE-08-2TPE	REUSE-09-2TPE	REUSE-10-2TPE	REUSE-11-2TPE

REUSE-11L	REUSE-12	REUSE-12L	REUSE-13
REUSE-11L-1HP	REUSE-12-1HP	REUSE-12L-1HP	REUSE-13-1HP
REUSE-11L-1MQ	REUSE-12-1MQ	REUSE-12L-1MQ	REUSE-13-1MQ
REUSE-11L-1TP	REUSE-12-1TP	REUSE-12L-1TP	REUSE-13-1TP
REUSE-11L-1TPE	REUSE-12-1TPE	REUSE-12L-1TPE	REUSE-13-1TPE
REUSE-11L-2BV	REUSE-12-2BV	REUSE-12L-2BV	REUSE-13-2BV
REUSE-11L-2BVC	REUSE-12-2BVC	REUSE-12L-2BVC	REUSE-13-2BVC
REUSE-11L-2MF	REUSE-12-2MF	REUSE-12L-2MF	REUSE-13-2MF
REUSE-11L-2MQ	REUSE-12-2MQ	REUSE-12L-2MQ	REUSE-13-2MQ
REUSE-11L-2TP	REUSE-12-2TP	REUSE-12L-2TP	REUSE-13-2TP
REUSE-11L-2TPE	REUSE-12-2TPE	REUSE-12L-2TPE	REUSE-13-2TPE

REUSE-ACC-MON	REUSE-FP-MON
REUSE-MLT-1	REUSE-PED-BV
REUSE-PED-MON	REUSE-MLT
REUSE-FP-BV	REUSE-PED-HAND
REUSE-MLT-2	REUSE-FP-HAND

GMDN Code and Term 34978 Blood Pressure Cuff, Reusable

UMDNS Code and Term 11072 Cuffs

Basic UDI-DI 0732094GMN901043EY



Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

Approval



Joshua Kim, Sr. Manager
Global Regulatory Affairs

2022-02-16

Date

Skaneateles Falls, NY USA

Place of Issue

Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

Appendix A: Standards and Common Specifications

Standards Applied	Number	Version/Date of Issue	Title
Regulation 2017/745	EN ISO 13485	2016	Medical Devices - Quality Management Systems - Requirements for Regulatory Purposes
	EN ISO 15223-1	2016	Medical Devices - Symbols to be Used with medical Device Labels, Labelling and Information to be Supplied - Part 1: General Requirements
	EN ISO 10993-1	2018	Biological evaluation of medical devices - Part 1: Evaluation and testing
	EN ISO 10993-5	2009	Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity
	EN ISO 10993-10	2013	Biological evaluation of medical devices — Part 10: Tests for irritation and skin sensitization
	EN ISO 81060-1	2012	Non-invasive sphygmomanometers - Part 1: Requirements and test methods for non- automated measurement type
	EN ISO 80369-5	2016	Small-bore connectors for liquids and gases in healthcare applications - Part 5: Connectors for limb cuff inflation applications
	EN 62366-1	2015	Medical Devices - Part 1: Application of Usability Engineering to Medical Devices



Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

Document Change History

<u>Version</u>	<u>Description</u>	<u>Author</u>	<u>Date</u>
<u>A</u>	Updated to SAP, Removed the neonate cuffs and FlexiPort Cuffs from DoC. Removed component parts that are not accessories and not supported in the technical file or compliant with MDD labeling.	S. Schmidt	2011-01-07
<u>B</u>	Updated to new format. Updated EU REP address, removed Notified Body since this is a class I product.	S. Schmidt	2011-08-12
<u>C</u>	Added annex for Saudi Arabia submission. Updated to new format. Removed bladders, since these are components and are not CE marked (all 5089-X numbers and 5200-04, 05, 06, and 11, checked pkg labels). Added 4500-01, 4500-200, 5082-21H, 5082-64, 5082-65, 5200-01, 5200-02, 5200-03, 5200-10 (checked all pkg labels include CE mark). Added 2-pc cuffs being transferred from Jungingen to Tijuana.	S. Schmidt	2012-01-10
<u>D</u>	1) Added EN/ISO 81060-1 2) Removed EN 1060-1 and EN 1060-2 since both are formally withdrawn as of May 31, 2015 and all listed cuffs have met EN/ISO 81060-1 (see 60066583) which is in force. 3) Corrected UMDNS code 11072 to identify device name and not definition 4) Operations provided listed "#" numbers that correlate to device numbers listed within test report 60066583. 5) Updated table in annex as a duplicate of the table within the DoC.	J. Strong	2015-05-21
<u>E</u>	Since cuffs (no bladder) are CE marked for the entire device (cuff & bladder), the following cuffs (no bladder) material numbers are being added back onto this DoC: 5082-01, 5082-01H, 5082-02, 5082-11, 5082-16, 5082-21, 5082-24. Since descriptions have been included in table in the main declaration of conformity, the table in the annex isn't needed, so that data has been removed.	M. McGovern	2016-01-08



Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.

F	Updated for MDR	C. Lefancheck	2020-03-23
G	Class I updates for MDR	C. Lefancheck	2021-06-15
H	DoC Correction & SRN	K. Love	2021-08-20
J	DoC Correction, New SAP Change Number	K. Love	2021-08-23
K	Updated to new template, added SRN for Skaneateles and updated the part number list to a table format.	K. Ockenfels	2021-10-12
L	Updated to new template, added Intended Purpose Statement	K. Love	2021-11-08
M	Updated part list to remove the -SC cuffs which are now classified as Kits due to small bore compliance	K Ockenfels	2022-02-04