

Declaration of Conformity

USS-048

We hereby declare, under our sole responsibility, that the devices specified below meet the relevant provisions of the Council Directive concerning medical devices- 93/42/EEC and the Essential Principles. This is also a declaration made in accordance with the requirements of Clause 1.8 of schedule 3 of the Australian Therapeutic Goods (Medical Devices) Regulations 2002 relating to the stated device.

| | |
|---|---|
| Issued by Manufacturer: | Covidien llc 15 Hampshire Street Mansfield, MA 02048, U.S.A. |
| Original Date/Place of Issue: | 08/04/1995 North Haven, CT |
| Type of Devices: | Skin/Fascia Staplers |
| Device Name: | See Attached |
| Product Category(ies) Listed on Current MDD Certificate: | Surgical Staple, Clip Products and Accessories |
| MDD Classification/Reorder Codes/GMDN Codes: | See Attached |
| Conformity Assessment: | Directive 93/42/EEC on Medical Devices (MDD), For Class IIa/IIb: Annex II excluding (4) For Class I sterile: Annex V All Class I, non-sterile, non-measurement devices listed on Declarations of Conformance are not regulated by TÜV SÜD P.S. and follow conformity assessment procedures set out in Annex VII. |
| EC Certificate: | G1 077608 0079 Rev 00 (expires 26-May-2024) G2S 077608 0072 Rev. 00 (expires 23-May-2024) |
| Certificate of Conformity Valid Until: | 23-May-2024 |
| Standards Associated: | See Attached |

Authorized Representative in EU

Covidien Ireland Limited
IDA Business and Technology Park
Tullamore, Ireland

Notified Body

TUV SUD Product Service GmbH
Ridlerstrasse 65,
80339 Munich, Germany (0123)

Revision Date: October 3, 2019

Page 1 of 5


Mary Mellows
Manager, Regulatory Affairs



Declaration of Conformity

USS-048

| <i>Reorder Code</i> | <i>Description</i> | <i>MDD Class</i> | <i>MDD Rule</i> | <i>GMDN Code</i> | <i>Date Added to Declaration MM/DD/YYYY</i> | <i>Reorder Code Status</i> |
|---------------------|---|------------------|-----------------|--|---|----------------------------|
| 050258 | SM™ Auto Suture™ Skin Stapler Reload 35 | IIb | 8 | Surgical Staple, non-bioabsorbable [35615] | 8/4/1995 | Current |
| 050284 | SM™ Auto Suture™ Skin Stapler Reload 35W | IIb | 8 | Surgical Staple, non-bioabsorbable [35615] | 8/4/1995 | Current |
| 054006 | Signet™ Auto Suture™ Skin Stapler 35W | IIa | 7 | Skin Stapler/staple, non-bioabsorbable [35884] | 8/4/1995 | Current |
| 054887 | Royal™ Auto Suture™ Skin Stapler 35W | IIa | 7 | Skin Stapler/staple, non-bioabsorbable [35884] | 8/4/1995 | Current |
| 059035 | Multifire Premium™ Auto Suture™ Pistol Grip Skin Stapler 35 | IIa | 7 | Skin Stapler/staple, non-bioabsorbable [35884] | 8/4/1995 | Current |
| 059036 | Multifire Premium™ Auto Suture™ Pistol Grip Skin Stapler Reload 35 | IIa | 7 | Skin Stapler/staple, non-bioabsorbable [35884] | 8/4/1995 | Current |
| 059037 | Multifire Premium™ Auto Suture™ Pistol Grip Skin Stapler 35W | IIa | 7 | Skin Stapler/staple, non-bioabsorbable [35884] | 8/4/1995 | Current |
| 059038 | Multifire Premium™ Auto Suture™ Pistol Grip Skin Stapler Reload 35W | IIa | 7 | Skin Stapler/staple, non-bioabsorbable | 8/4/1995 | Current |

Revision Date: October 3, 2019
Page 2 of 5


Mary Mellows
Manager, Regulatory Affairs




Declaration of Conformity

USS-048

| [35884] | | | | | | | |
|------------|---|---------------|---|--|-----------|---------|--|
| 060210 | SFS™ Auto Suture™ Fascia Stapler 20W | I non-sterile | 1 | Surgical stapler, reusable [32369] | 8/4/1995 | Current | |
| 070614 | DFS™ Auto Suture™ Fascia Stapler 20W | IIb | 8 | Open-surgery manual linear stapler, single use [59873] | 5/23/2019 | Current | |
| 150462 | Auto Suture™ Premium Extractor | I sterile | 1 | Surgical staple remover, single-use [47192] | 8/4/1995 | Current | |
| 8886803512 | Appose™ ULC Auto Suture™ Slim Body Skin Stapler 35 | IIa | 7 | Skin Stapler/staple, non-bioabsorbable [35884] | 2/16/2000 | Current | |
| 886803712 | Appose™ ULC Auto Suture™ Slim Body Skin Stapler 35W | IIa | 7 | Skin Stapler/staple, non-bioabsorbable [35884] | 2/16/2000 | Current | |


Mary Mellows
Manager, Regulatory Affairs



Standards/Directives List

| Standard/Directive | Year | Title |
|---------------------|-------------|--|
| EN 556-1 + AC | 2001 + 2006 | Sterilization of Medical Devices - Requirements for Medical Devices to Be Designated "Sterile" - Part 1: Requirements for Terminally Sterilized Medical Devices |
| EN ISO 11135-1 | 2007 | Sterilization of health care products — Ethylene oxide — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices |
| EN ISO 11737-1 + AC | 2006 + 2013 | Sterilization of medical devices - Microbiological methods - Part 1: Determination of a population of microorganisms on products. |
| EN ISO 11137-1 + AC | 2006 + 2013 | Sterilization of health care products - Radiation - Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices. |
| EN ISO 11137-2 | 2013 | Sterilization of health care products - Radiation - Part 2: Establishing the sterilization dose |
| EN ISO 11607-1 + A1 | 2010 + 2014 | Packaging for terminally sterilized medical devices Part 1: Requirements for materials, sterile barrier systems and packaging systems |
| EN ISO 11607-2 + A1 | 2006 + 2014 | Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes |
| ISO 15223-1 | 2012 | Medical devices - Symbols to be used with medical device labels, labeling and information to be supplied - Part 1: General requirements |
| EN ISO 13485 + AC | 2012 + 2012 | Medical devices. Quality management systems. Requirements for regulatory purposes. |
| EN ISO 14630 | 2012 | Non-active surgical implants – General Requirements |
| EN 1041 | 2008 | Information supplied by the manufacturer with medical devices. |
| EN ISO 14971 | 2012 | Medical devices -- Application of risk management to medical devices. |
| EN ISO 17664 | 2004 | Sterilization of medical devices – Information to be provided by the manufacturer for the processing of resterilizable medical devices. |
| EN 62366 | 2014 | Medical devices — Application of usability engineering to medical devices |
| ISO 14644-1 | 2015 | Cleanrooms and associated controlled environments - Part 1: Classification of air cleanliness by particle concentration |
| ISO 14644-2 | 2015 | Cleanrooms and associated controlled environments - Part 2: Monitoring to provide evidence or cleanroom performance related to air cleanliness by particle concentration |

| | | |
|---------------------|-------------|--|
| ISO 14644-3 | 2005 | Cleanrooms and associated controlled environments - Part 3: Test methods |
| EN ISO 10993-1 + AC | 2009 + 2010 | Biological evaluation of medical devices - Part 1: Evaluation and testing. |
| EN ISO 10993-3 | 2014 | Biological evaluation of medical devices - Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity |
| EN ISO 10993-4 | 2009 | Biological evaluation of medical devices - Part 4: Selection of tests for interactions with blood |
| EN ISO 10993-5 | 2009 | Biological evaluation of medical devices Part 5: Tests for in vitro cytotoxicity |
| EN ISO 10993-6 | 2009 | Biological evaluation of medical devices Part 6: Tests for local effects after implantation |
| EN ISO 10993-7 + AC | 2008 + 2009 | Biological evaluation of medical devices. Part 7: Ethylene oxide sterilization residuals |
| ISO 10993-10 | 2010 | Biological evaluation of medical devices Part 10: Tests for irritation and skin sensitization |
| EN ISO 10993-11 | 2009 | Biological evaluation of medical devices Part 11: Tests for systemic toxicity |